



Outcome on EPIZONE Extension on VER/ VNN: Diagnostics, proficiency test and qRT-PCR validation

Bigarré, Laurent; Panzarin, Valentina; Baud, Marine; Jonstrup, Søren Peter; Jansson, Eva ; Isaksson, Mats ; Engelsma, Marc ; Olesen, Niels Jørgen; Bovo, Giuseppe

Published in:
16th Annual Meeting of the National Reference Laboratories for Fish Diseases

Publication date:
2012

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Bigarré, L., Panzarin, V., Baud, M., Jonstrup, S. P., Jansson, E., Isaksson, M., Engelsma, M., Olesen, N. J., & Bovo, G. (2012). Outcome on EPIZONE Extension on VER/ VNN: Diagnostics, proficiency test and qRT-PCR validation. In *16th Annual Meeting of the National Reference Laboratories for Fish Diseases*

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

OUTCOME ON EPIZONE EXTENSION ON VER/ VNN: Diagnostics, proficiency test and qRT-PCR validation

¹ Laurent Bigarré, ² Valentina Panzarin, ¹ Marine Baud, ³ Soren Peter Jonstrup, ⁴ Eva Jansson, ⁴ Mats Isaksson, ⁵ Marc Engelsma, ³ Niels Jørgen Olesen, ² Giuseppe Bovo

¹ Agence nationale de la sécurité sanitaire, alimentation, environnement et travail, ANSES, France. ² Istituto Zooprofilattico Sperimentale delle Venezie, IZSVE, Italy. ³ Technical University of Denmark, National Veterinary Institute, DTU VET, Denmark. ⁴ Statens Veterinärmedicinska Anstalt, SVA, Sweden. ⁵ Central Veterinary Institute of Wageningen UR, CVI, The Netherlands

Introduction

Betanodaviruses are genetically very diverse with at least five genogroups described and a large range of variants within each group, including reassortants carrying components of two groups (Panzarin *et al.*, 2011; Toffolo *et al.*, 2007). Their genome is composed of two strands of RNA: RNA1 and RNA2. Till now, only one universal real-time RT-PCR method has been published which detects all variants of RNA2 (Panzarin *et al.*, 2010) in clinical cases. An additional method based on detection of RNA1 would be useful to confirm results obtained from RNA2. Another advantage of detecting RNA1 is to bring valuable genetic information on the second genome component of a given isolate. ANSES has developed new DNA probes targeting RNA1 with the goal of detecting all genotypes of nodaviruses. The aim of the project is thus:

- To organize, conduct and report an inter-laboratory proficiency test for detection of aquatic nodaviruses by real time RT-PCR targeting RNA1 and RNA2, respectively.
- To test a newly developed real-time RT-PCR targeting RNA1: sensitivity, specificity, range of detection and genetic information provided by sequencing the PCR product.

Materials & methods

Primers and probes specific for RNA 1 and RNA2 were ordered by ANSES and aliquots from the same batch were provided to all partners. IZSVE produced for all partners 10 samples (1-11) of inactivated isolates produced in cell culture and covering 4 genogroups. The samples were forwarded by ANSES and RNA had to be extracted by each partner. In the meantime, ANSES produced and distributed RNA extracted from healthy or infected fish (2 genogroups), or cell culture; one sample from cell culture had to be serially diluted to test the sensitivity of each method in partners' hands. Samples were tested in duplicates and the mean Ct values reported.

Results

A total of 192 virus-containing samples were tested (negative controls excluded), implicating 5 partners performing each two or four methods (16 tubes with virus * 12 methods). For each genetic component (RNA1 or RNA2), the same ratio of detection (80 / 96) was found, indicating that both RNA1 and RNA2 probes are globally equivalent for detection of all genotypes. When the samples sent were considered, the rate of false positives was 5,5% (2/36) and the rate of false negative 13,5% (31/228). Both RNA1 and RNA2 proved to be valuable targets for viral detection, with advantages and disadvantages depending on the method and the viral isolates.

Conclusions

Both sets of the nodavirus probes proved to be useful for detection of betanodaviruses in each lab's usual conditions. Ideally, both RNA should be targeted for samples exhibiting high Ct values (for instance in healthy carriers) to confirm the presence of low loads of virus.

References

- Toffolo, V., Negrisol, E., Maltese, C., Bovo, G., Belvedere, P., Colombo, L. & Dalla Valle, L. (2007). Phylogeny of betanodaviruses and molecular evolution of their RNA polymerase and coat proteins. *Molecular Phylogenetics and Evolution* 43, 298-308.
- Panzarin, V., Patarnello, P., Mori, A., Rampazzo, E., Cappelozza, E., Bovo, G. & Cattoli, G. (2010). Development and validation of a real-time TaqMan PCR assay for the detection of betanodavirus in clinical specimens. *Arch Virol* 155, 1193-1203.
- Panzarin, V., Fusaro, A., Monne, I., Cappelozza, E., Patarnello, P., Bovo, G., Capua, I., Holmes, E. C. & Cattoli, G. (2011). Molecular epidemiology and evolutionary dynamics of betanodavirus in southern Europe. *Infect Genet Evol.*